EUROPEAN UNION RECOGNISED ORGANISATION (EU RO) MUTUAL RECOGNITION TYPE APPROVAL CERTIFICATE

Certificate No:
MRA000003F

In accordance with Article 10.1 of EU Regulation 391/2009

This Certificate is issued to

ROTA YOKOGAWA GmbH & Co. KG

Wehr, Baden-Württemberg, **Germany**

for
Flow Gauges/Transmitters

with type designation(s)
Rotamass Total Insight

The product is found to comply with
EU RO Mutual Recognition Technical Requirements for Flow Gauges/Transmitters

Intended service
Coriolis mass flow and density meters intended for a wide range of use. See product description on page 2.

Temperature [°C]:	**5°C to 70°C**
Vibration:	**Sensor: ±1.6 mm / 4.0 g, Transmitter: ±1.0 mm / 0.7g**
EMC:	**General power distribution zone**
IP Code:	**IP66/67**

This is to certify:

that the Product referred to herein has been inspected for the Manufacturer, pursuant to the relevant requirements of the European Union Recognised Organisation Mutual Recognition procedure, required by Article 10.1 of EU Regulation 391/2009, and has been found in accordance with those requirements.

This Certificate is valid until **2026-09-14**.

Issued at **Høvik** on **2021-09-15**
DNV local station: **Augsburg**

for **DNV**

Approval Engineer: **Ståle Sneen**

Marta Alonso Pontes
Head of Section

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid.
The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.
LEGAL DISCLAIMER: Unless otherwise stated in the applicable contract with the holder of this document, or following from mandatory law, the liability of DNV AS, its parent companies and their subsidiaries as well as their officers, directors and employees ("DNV") arising from or in connection with the services rendered for the purpose of the issuance of this document or reliance thereon, whether in contract or in tort (including negligence), shall be limited to direct losses and under any circumstance be limited to 300,000 USD.

Product description

Coriolis mass flow and density meter Rotamass Total Insight
Variants covered by the Type Approval are specified by model code[1)] containing at least:

Transmitter:	Rotamass Total Insight Transmitter[2)]
Sensor:	Nano, Supreme, Giga, Prime, Intense or Hygienic
Meter size[1)]:	10, 15, 20, 25, 34, 36, 38, 39, 40, 50, 80, 1H or 1F
Design and housing[1)]:	Remote installation types
Communication type and I/O[1)]:	JA, JB, JC, JD, JE, JF, JG, JH, JJ, JK, JL, JM or JN (HART); M0, M2, M3, M4, M5, M6 or M7 (MODBUS).

[1)] For further details and description of codes refer to effective General Specifications of Rotamass Total Insight.
[2)] In combination with any approved sensor/detector.

For Hardware and Firmware versions, please see the applicable Software Instruction Manual.
Accuracy class depending on the chosen model, please see the applicable General Specifications.

Manufactured by

ROTA YOKOGAWA GmbH & Co. KG
Rheinstraße 8, 79664 Wehr, Germany

Application/Limitation

The flow meters are intended for use in control and monitoring systems subject to classification.

The wetted parts (sensors) have been tested and approved according to the general requirements of IACS UR E10. Application of wetted parts must be made considering properties such as temperature, pressure, and corrosive properties of the measured medium, per the relevant requirements of the classing RO.

Type Approval documentation

Marking of product

The products to be marked with:
- manufacturer name
- model name
- serial number and date of manufacture
- particulars or ratings for transmitter (including IP degree and class of accuracy)

Other conditions

The FLOW GAUGES/TRANSMITTERS have been verified for compliance with EU Mutual Recognition Technical Requirements for FLOW GAUGES/TRANSMITTERS version 0.0. The testing covers IACS UR E10 Rev.7 / Rev.8.
Relevant testing according to IACS UR P2.11:
- Fire endurance test was not found relevant as the device is not shipped with gaskets or seals for the flanges
- Pressure test and tightness test are routinely performed during production

Environmental test parameters		DNV location classes
Temperature:	5°C and 70°C	B
Vibration:	Sensor: ±1.6 mm / 4.0 g, Transmitter: ±1.0 mm / 0.7 g	Sensor: B, Transmitter: A
Humidity:	95%RH @ 55°C, damp heat cyclic	B
EMC:	General power distribution zone	A
Enclosure:	IP66/67 (according to IEC 60529)	B

Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate

Periodical assessment will be performed annually and at renewal of the certificate.

Generic Statement for EU RO MR Type Approval Certificate

When a product is presented with this EU RO MR Type Approval Certificate for given application, its acceptability with regards to the limitations stated in the certificate conditions defined in 1b, 1c and 1d of the applied Technical Requirement will be evaluated by the EU RO in charge of classing the ship or being in charge of the unit/system certification.

In accordance with Article 10 of Regulation (EC) No 391/2009 of the European Parliament and of the Council of 23 April 2009 "on common rules and standards for ship inspection and survey organizations", the following organizations, recognized by the EU on this date, have agreed on the technical and procedural conditions under which they will mutually recognize this certificate:

- American Bureau of Shipping (ABS);
- Bureau Veritas (BV);
- China Classification Society (CCS);
- Croatian Register of Shipping (CRS);
- DNV;
- Indian Register of Shipping (IRS);
- Korean Register (KR);
- Lloyd's Register Group Ltd. (LR);
- Nippon Kaiji Kyokai General Incorporated Foundation (ClassNK);
- Polish Register of Shipping (PRS);
- RINA Services S.p.A. (RINA);
- Russian Maritime Register of Shipping (RS).

The scheme for the mutual recognition of class certificates for materials, equipment and components laid down by Article 10(1) of Regulation (EC) No 391/2009 is only enforceable within the Union in respect of ships flying the flag of a Member State. As far as foreign vessels are concerned, the acceptance of relevant certificates remains at the discretion of relevant non-EU flag States in the exercise of their exclusive jurisdiction, notably under the United Nations Convention on the Law of the Sea (UNCLOS). (In accordance with COMMISSION IMPLEMENTING REGULATION (EU) No 1355/2014 amending Regulation (EC) No 391/2009 - recital (25)).