DNV·GL

Certificate No:
TAA00001JW

TYPE APPROVAL CERTIFICATE

This is to certify:

That the Temperature Sensor

with type designation(s)
RTD Temperature Probe 902020/99-3802003-1-5-104 / 000

Issued to
JUMO GmbH & Co. KG
Fulda, Hessen, Germany

is found to comply with
DNV GL rules for classification – Ships, offshore units, and high speed and light craft

Application :

Product(s) approved by this certificate is/are accepted for installation on all vessels classed by DNV GL.

Temperature	**B**
Humidity	**B**
Vibration	**A**
EMC	**B***
Enclosure	**B (IP65)**

***) Due to the nature of the equipment (passive resistors) EMC testing has not been performed.**

Issued at **Hamburg** on **2017-12-14**
This Certificate is valid until **2022-12-13**.
DNV GL local station: **Magdeburg**

Approval Engineer: **Jens Dietrich**

for **DNV GL**

Joannis Papanuskas
Head of Section

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid.
The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.

© DNV GL 2014. DNV GL and the Horizon Graphic are trademarks of DNV GL AS.

Job Id: **262.1-025659-1**
Certificate No: **TAA00001JW**

Product description

RTD Temperature Probe 902020/99-3802003-1-5-104 / 000 Srew-In Temperature Probes
Operating temperature range:
-40°C ... +100°C terminal head
-50°C ... +200°C protection tube
Response times:
water@0.4m/s: t0.5 < 2.0s, t0.9 < 4.0s
air@3m/s: t0.5 < 25s, t0.9 < 70s
Measuring element: 2x Pt100 DIN EN 60751, Kl.B
Cable screw connection: M20x1.5
Terminal head type: BA, IP65
Protection tube: Stainless steel 1.4404

Application/Limitation

The Type Approval covers hardware listed under Product description. When the hardware is used in applications to be classed by DNV GL, documentation for the actual application is to be submitted for approval by the manufacturer of the application system in each case. Reference is made to DNV GL RU SHIP Pt. 4 Ch. 9 Sec. 1.

Tests carried out

Applicable Tests according to Class Guidleine DNV GL CG-0339, November 2016

Marking of product

Maker, type, serial-number

Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer’s product type marking and the type approval certificate

Periodical assessment is to be performed after 2 years and after 3.5 years. A renewal assessment will be performed at renewal of this certificate.

END OF CERTIFICATE