

TYPE APPROVAL CERTIFICATE

This is to certify:**That the Measurement Converter**with type designation(s)
VEGAMET 624, 625, VEGASCAN 693

Issued to

VEGA Grieshaber KG
Schiltach, Germany

is found to comply with

DNV GL rules for classification – Ships, offshore units, and high speed and light craft**Application :****Product(s) approved by this certificate is/are accepted for installation on all vessels classed by DNV GL.****Location classes:**

Temperature	B
Humidity	B
Vibration	A
EMC	B
Enclosure	A

Issued at **Hamburg** on **2020-06-05**This Certificate is valid until **2025-06-04**.for **DNV GL**DNV GL local station: **Augsburg**Approval Engineer: **Dariusz Lesniewski**

Joannis Papanuskas
Head of Section

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid. The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.

LEGAL DISCLAIMER: Unless otherwise stated in the applicable contract with the holder of this document, or following from mandatory law, the liability of DNV GL AS, its parent companies and subsidiaries as well as their officers, directors and employees ("DNV GL") arising from or in connection with the services rendered for the purpose of the issuance of this document or reliance thereon, whether in contract or in tort (including negligence), shall be limited to direct losses and under any circumstance be limited to 300,000 USD.



Product description

Signal conditioning instrument, series VEGAMET 624 Type MET624...
VEGAMET 625 Type MET625...
VEGASCAN 693 Type SCAN693...

Main characteristics

Power supply	24V DC rated (range 20...250V DC)
Signal inputs	MET624: 1x 4...20mA / HART, 2-wire, sensor power supply MET625: 2x HART, 2-wire, sensor power supply SCAN693: 15x (Ex-version 5x) HART, 2-wire, sensor power supply (shielded cable for HART communication to be used)
Outputs	1x signal fault relay 3x current output 3x operating relay (for MET624/625 only)
Communication interface (optional)	RS232- or Ethernet-Interface
Graphic LC-display	50x25mm with background lighting
Indication LEDs	2x green, 3x yellow, 1x red
Adjustment	by 4x keys in front for menu adjustment by PACTware with VEGA-DTM
Mounting	wall mounting with adapter plate
Degree of protection	IP30 (instrument) / IP20 (plug-in socket)
Software version	V1.xx
Ex-proof protection	as the certificate TÜV 03 ATEX 2269

Application/Limitation

EMC in the range 2 GHz to 6 GHz according to DNVGL-CG-0339, December 2019 has not been verified. EMC up to 6 GHz must additionally be documented for installation on ships contracted for construction on or after 2022-01-01.

Type Approval documentation

Tests carried out

Applicable tests according to class guideline DNVGL-CG-0339, December 2019.

Marking of product

The products to be marked with:

- manufacturer name
- model name
- serial number
- power supply ratings

Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate



Job Id: **262.1-033352-1**
Certificate No: **TAA00002T6**

Periodical assessment is to be performed after 2 years and after 3.5 years. A renewal assessment will be performed at renewal of the certificate.

END OF CERTIFICATE