



Certificate no.:
TAA00003RN

TYPE APPROVAL CERTIFICATE

This is to certify:

that the **Portable Multi-Gas Detector**

with type designation(s)
X-am 2800, X-am 5800

issued to

Dräger Safety AG & Co. KGaA
Lübeck, Germany

is found to comply with

DNV rules for classification – Ships, offshore units, and high speed and light craft

Application:

Product(s) approved by this certificate is/are accepted for installation on all vessels classed by DNV.

Location classes:

Temperature	C
Humidity	B
Vibration	n/a
EMC	B
Enclosure	B, C, D (*see Product description)

Issued at **Hamburg** on **2025-11-26**

This Certificate is valid until **2030-11-25**.

DNV local unit: **Hamburg – CMC North/East**

for **DNV**

Approval Engineer: **Dariusz Lesniewski**

.....

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid.
The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.

LEGAL DISCLAIMER: Unless otherwise stated in the applicable contract with the holder of this document, or following from mandatory law, the liability of DNV AS, its parent companies and their subsidiaries as well as their officers, directors and employees ("DNV") arising from or in connection with the services rendered for the purpose of the issuance of this document or reliance thereon, whether in contract or in tort (including negligence), shall be limited to direct losses and under any circumstance be limited to USD 300 000.

Product description

Dräger X-am 2800/ 5800 portable Multi-Gas Detector for the simultaneously detection of flammable and toxic gases and/or vapours and oxygen in ambient air.

Features (main)	Dräger X-am 2800 (MQG 0200)	Dräger X-am 5800 (MQG 0210)
Number of measured gases	1 to 4	1 to 6
Catalytic sensor (XDS 03**)	Yes	Yes
PID neo	No	Yes
X-am Pump (GEP 02**) (DNV certificate TAA0000144)	Yes	Yes

Technical data (main)

Display	High-resolution matrix display
Alarms	Visual 'red' (gas alarms) 'yellow' (device alarms) Audible Multi-tone Vibration
Power supply	LiMH battery, replaceable, rechargeable
Pump operation	Optional
Ex-certification	IECEX
*Location class 'Enclosure' / Degree of protection	B = (protected areas) C = IP 56 / IP 66 with pump (open deck) D = IP 68 (submerged)

Software version: 1.x.x

Approval conditions

The Type Approval covers hardware listed under Product description. When the hardware is used in applications to be classed by DNV, documentation for the actual application is to be submitted for approval by the manufacturer of the application system in each case. Reference is made to DNV rules for classification of ships Pt.4 Ch.9 Control and monitoring systems.

Application/Limitation

For use in hazardous areas provisions listed in the valid Ex-certificates to be observed.

Type Approval documentation

Tests carried out

Applicable tests according to class guideline DNV-CG-0339, August 2021.
ISO 19891-1:2017

Marking of product

The products to be marked with:

- manufacturer name
- model name
- serial number
- power supply ratings

Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate

Periodical assessment is to be performed after 2 years and after 3.5 years. A renewal assessment will be performed at renewal of the certificate.

END OF CERTIFICATE